



P.O. Box 86 | Wooster, OH 44691  
Phone: 330.264.5678 | Fax: 330.264.2944  
[www.akronbrass.com](http://www.akronbrass.com)

## **Calibration of the Magmeter in the AkroFlow™ Portable Flow Meter**

During the manufacturing process the magmeter is wet calibrated on a test bench which is traceable to NIST standards.

After calibration, the accuracy of every AkroFlow meter will be within:  
+/- 1% of reading (between 10% and 100% of the maximum flow rate)